

(+48) 32 453 91 70

⊠ info@pol-eko.com.pl

www.pol-eko.com.pl

POL-EKO A.Polok-Kowalska sp.k. 44-300 Wodzisław Śląski ul. Kokoszycka 172 C

Data sheet

Stationary sampler PP 2002+E/1



The photo above is for reference only, may show additional options not included in standard equipment. The real appearance, particularly color and structure of the material may differ from the ones presented in the photo.





(+48) 32 453 91 70info@pol-eko.com.plwww.pol-eko.com.pl


POL-EKO A.Polok-Kowalska sp.k. 44-300 Wodzisław Śląski ul. Kokoszycka 172 C

Sampling system	peristaltic pump
Sample storing	stable temperature +4°C regardless of ambient conditions
Menu language	EN, FR, PL, CZ, RO, LT, IT
Medium	liquids of conductivity min 20 $\mu S/cm$ and max temp. $60^{\circ} C$
Hose blowing	before and after sample taking
Sampling mode	automatic time proportional, flow proportional, event or manual
Sampling height [m]	max 8
Sample volume [ml]	regulated 109990
Hose lenght [m]	8 standard
Hose diameter [mm]	12/13
Distributor	round
Number of bottles x capacity [I]	1 x 25
Housing	acid-proof stainless steel with 40 mm insulation
Ambient temperature [°C]	-20+45
Operating conditions	indoor or outdoor installation
Controller	microprocessor, graphic display with contrast control, bottle filling overview
Programming	5 programs, 8 tasks each
Data logging	SD card + Sampler Viewer software
Communication	RS 232 or RS 485
Input signals	8 analogue, 4 binary
Output signals	4 binary

overall	dims	[mm]
Ovciun	uiiii	[

Width A	630
Height B	1070
Depth C	660





weight [kg] 90

230 V 50 Hz
450
24 months
POL-EKO

